



stngco.com

CORPORATE OFFICE

2691 Saturn Street, Brea, CA 92821
Tel: (714) 524-0663, Fax: (714) 364-8113



ST&G CORPORATION

Rm. 911 43, Yangsan-ro
Yeongdeungpo-gu, Seoul, Korea



MDSS GmbH

Schiffgraben 41, 30175 Hannover, Germany



REF: 3231T

**SACH Foot Adapter for 375lbs,
Ti**

Weight: 88 grams
Weight Limit: 170kg / 375lbs
Torque Recommendation:
Attachment bolt cap screw: 30 Nm



3231T IFU REV. D (07-12-23)



stngco.com

CORPORATE OFFICE

2691 Saturn Street, Brea, CA 92821
Tel: (714) 524-0663, Fax: (714) 364-8113



ST&G CORPORATION

Rm. 911 43, Yangsan-ro
Yeongdeungpo-gu, Seoul, Korea



MDSS GmbH

Schiffgraben 41, 30175 Hannover, Germany



REF: 3231T

**SACH Foot Adapter for 375lbs,
Ti**

Weight: 88 grams
Weight Limit: 170kg / 375lbs
Torque Recommendation:
Attachment bolt cap screw: 30 Nm



3231T IFU REV. D (07-12-23)



stngco.com

CORPORATE OFFICE

2691 Saturn Street, Brea, CA 92821
Tel: (714) 524-0663, Fax: (714) 364-8113



ST&G CORPORATION

Rm. 911 43, Yangsan-ro
Yeongdeungpo-gu, Seoul, Korea



MDSS GmbH

Schiffgraben 41, 30175 Hannover, Germany



REF: 3231T

**SACH Foot Adapter for 375lbs,
Ti**

Weight: 88 grams
Weight Limit: 170kg / 375lbs
Torque Recommendation:
Attachment bolt cap screw: 30 Nm



3231T IFU REV. D (07-12-23)



stngco.com

CORPORATE OFFICE

2691 Saturn Street, Brea, CA 92821
Tel: (714) 524-0663, Fax: (714) 364-8113



ST&G CORPORATION

Rm. 911 43, Yangsan-ro
Yeongdeungpo-gu, Seoul, Korea



MDSS GmbH

Schiffgraben 41, 30175 Hannover, Germany



REF: 3231T

**SACH Foot Adapter for 375lbs,
Ti**

Weight: 88 grams
Weight Limit: 170kg / 375lbs
Torque Recommendation:
Attachment bolt cap screw: 30 Nm



3231T IFU REV. D (07-12-23)

1. Application
The 3231T, SACH Foot Adapter for 375lb, TI is to be exclusively used for the prosthetic fitting of lower extremity amputations.

2. Description
The SACH Foot Adapter for 375lb, TI is used to connect to a foot with single bolt attachment hole. The pyramid is connected to a pyramid receiver connection for anchoring and alignment setting. Apply Locate to the screw once the alignment and connection is finalized.

3. Warranty
Warranty applies only when the product is used as intended without modification, following all manufacturer's recommendations. ST&G offers 2 years warranty.

4. Declaration of Conformity
ST&G as manufacturer declares that The 3231T, SACH Foot Adapter for 375lb, TI conforms to the requirements of Council EU 2017/745 (MDR).

ST&G products are manufactured and tested for a particular weight and activity impact level. Use by another user for whom it was not originally manufactured may cause injury and voids any written or implied warranty. The liability shall fall upon those that resale/provide the product to anyone other than the original user delivered to. Not to be resold by private party.

Notice

Tighten all the screws, which are side by side, using the torque wrench



3

1. Application
The 3231T, SACH Foot Adapter for 375lb, TI is to be exclusively used for the prosthetic fitting of lower extremity amputations.

2. Description
The SACH Foot Adapter for 375lb, TI is used to connect to a foot with single bolt attachment hole. The pyramid is connected to a pyramid receiver connection for anchoring and alignment setting. Apply Locate to the screw once the alignment and connection is finalized.

3. Warranty
Warranty applies only when the product is used as intended without modification, following all manufacturer's recommendations. ST&G offers 2 years warranty.

2



4. Declaration of Conformity
ST&G as manufacturer declares that The 3231T, SACH Foot Adapter for 375lb, TI conforms to the requirements of Council EU 2017/745 (MDR).

ST&G products are manufactured and tested for a particular weight and activity impact level. Use by another user for whom it was not originally manufactured may cause injury and voids any written or implied warranty. The liability shall fall upon those that resale/provide the product to anyone other than the original user delivered to. Not to be resold by private party.

Notice

Tighten all the screws, which are side by side, using the torque wrench

3

1. Application
The 3231T, SACH Foot Adapter for 375lb, TI is to be exclusively used for the prosthetic fitting of lower extremity amputations.

2. Description
The SACH Foot Adapter for 375lb, TI is used to connect to a foot with single bolt attachment hole. The pyramid is connected to a pyramid receiver connection for anchoring and alignment setting. Apply Locate to the screw once the alignment and connection is finalized.

3. Warranty
Warranty applies only when the product is used as intended without modification, following all manufacturer's recommendations. ST&G offers 2 years warranty.

4. Declaration of Conformity
ST&G as manufacturer declares that The 3231T, SACH Foot Adapter for 375lb, TI conforms to the requirements of Council EU 2017/745 (MDR).

ST&G products are manufactured and tested for a particular weight and activity impact level. Use by another user for whom it was not originally manufactured may cause injury and voids any written or implied warranty. The liability shall fall upon those that resale/provide the product to anyone other than the original user delivered to. Not to be resold by private party.

Notice

Tighten all the screws, which are side by side, using the torque wrench



3

1. Application
The 3231T, SACH Foot Adapter for 375lb, TI is to be exclusively used for the prosthetic fitting of lower extremity amputations.

2. Description
The SACH Foot Adapter for 375lb, TI is used to connect to a foot with single bolt attachment hole. The pyramid is connected to a pyramid receiver connection for anchoring and alignment setting. Apply Locate to the screw once the alignment and connection is finalized.

3. Warranty
Warranty applies only when the product is used as intended without modification, following all manufacturer's recommendations. ST&G offers 2 years warranty.

2



4. Declaration of Conformity
ST&G as manufacturer declares that The 3231T, SACH Foot Adapter for 375lb, TI conforms to the requirements of Council EU 2017/745 (MDR).

ST&G products are manufactured and tested for a particular weight and activity impact level. Use by another user for whom it was not originally manufactured may cause injury and voids any written or implied warranty. The liability shall fall upon those that resale/provide the product to anyone other than the original user delivered to. Not to be resold by private party.

Notice

Tighten all the screws, which are side by side, using the torque wrench

3