



stngco.com

CORPORATE OFFICE

2691 Saturn Street, Brea, CA 92821
Tel: (714) 524-0663, Fax: (714) 364-8113



ST&G CORPORATION

Rm. 911 43, Yangsan-ro
Yeongdeungpo-gu, Seoul, Korea



MDSS GmbH

Schiffgraben 41, 30175 Hannover, Germany



REF: P211A

Pediatric Tube Clamp Assembly 22mm, Al

Weight: 133 grams

Weight Limit: 55 kg/ 122 lbs

Torque recommendation:

Alignment adjustment set screw – 10 Nm



P211A IFU REV D (5-1-25)



stngco.com


CORPORATE OFFICE

2691 Saturn Street, Brea, CA 92821
Tel: (714) 524-0663, Fax: (714) 364-8113



ST&G CORPORATION

Rm. 911 43, Yangsan-ro
Yeongdeungpo-gu, Seoul, Korea



MDSS GmbH

Schiffgraben 41, 30175 Hannover, Germany



REF: P211A

Pediatric Tube Clamp Assembly 22mm, Al

Weight: 133 grams

Weight Limit: 55 kg/ 122 lbs

Torque recommendation:

Alignment adjustment set screw – 10 Nm



P211A IFU REV D (5-1-25)



stngco.com

CORPORATE OFFICE

2691 Saturn Street, Brea, CA 92821
Tel: (714) 524-0663, Fax: (714) 364-8113



ST&G CORPORATION

Rm. 911 43, Yangsan-ro
Yeongdeungpo-gu, Seoul, Korea



MDSS GmbH

Schiffgraben 41, 30175 Hannover, Germany



REF: P211A

Pediatric Tube Clamp Assembly 22mm, Al

Weight: 133 grams

Weight Limit: 55 kg/ 122 lbs

Torque recommendation:

Alignment adjustment set screw – 10 Nm



P211A IFU REV D (5-1-25)



stngco.com


CORPORATE OFFICE

2691 Saturn Street, Brea, CA 92821
Tel: (714) 524-0663, Fax: (714) 364-8113



ST&G CORPORATION

Rm. 911 43, Yangsan-ro
Yeongdeungpo-gu, Seoul, Korea



MDSS GmbH

Schiffgraben 41, 30175 Hannover, Germany



REF: P211A

Pediatric Tube Clamp Assembly 22mm, Al

Weight: 133 grams

Weight Limit: 55 kg/ 122 lbs

Torque recommendation:

Alignment adjustment set screw – 10 Nm



P211A IFU REV D (5-1-25)

1. Application
The P211A,Pediatric Tube Clamp Assembly , 22mm is to be exclusively used for the prosthetic fitting of lower extremity amputations.

2. Description
The Tube Clamp Assembly, 22mm is used to connect any component with 22mm tube system. The four set screws are used for anchoring the mating pyramid system as well as for making angular adjustment for alignment. Apply recommended torque to alignment adjustment screws for connection. Apply locate to all screws once the alignment and connection is finalized.

3. Warranty
Warranty applies only when the product is used as intended without modification, following all manufacturer's recommendations. ST&G offers a 2 year warranty on this product.

4. Environment:
Store at room temperature, and avoid extremely hot, cold, or moist storage environments

Storage Temperature Range: between temperatures of -20°C to 80°C [-4°F and 176°F]

Do not store in wet or very hot environment



2

5. Declaration of Conformity
ST&G as manufacturer declares that the The P211A,Pediatric Tube Clamp Assembly , 22mm conforms to the requirements of Council EU 2017/745 (EU), ST&G products are manufactured and tested for a particular weight and activity impact level. Use by another user for whom it was not originally manufactured may cause injury and voids any written or implied warranty, or liability.

6. Useful Life
Service life of the product is covered by the warranty period.

7. Disposal
The device and its packaging must be disposed of in accordance with respective national/local environmental regulations.

Notice

Tighten all the screws, which are side by side, using the torque wrench



3

1. Application
The P211A,Pediatric Tube Clamp Assembly , 22mm is to be exclusively used for the prosthetic fitting of lower extremity amputations.

2. Description
The Tube Clamp Assembly, 22mm is used to connect any component with 22mm tube system. The four set screws are used for anchoring the mating pyramid system as well as for making angular adjustment for alignment. Apply recommended torque to alignment adjustment screws for connection. Apply locate to all screws once the alignment and connection is finalized.

3. Warranty
Warranty applies only when the product is used as intended without modification, following all manufacturer's recommendations. ST&G offers a 2 year warranty on this product.

Store at room temperature, and avoid extremely hot, cold, or moist storage environments

Storage Temperature Range: between temperatures of -20°C to 80°C [-4°F and 176°F]

Do not store in wet or very hot environment



2

5. Declaration of Conformity
ST&G as manufacturer declares that the The P211A,Pediatric Tube Clamp Assembly , 22mm conforms to the requirements of Council EU 2017/745 (EU), ST&G products are manufactured and tested for a particular weight and activity impact level. Use by another user for whom it was not originally manufactured may cause injury and voids any written or implied warranty, or liability.

6. Useful Life
Service life of the product is covered by the warranty period.

7. Disposal
The device and its packaging must be disposed of in accordance with respective national/local environmental regulations.

Notice

Tighten all the screws, which are side by side, using the torque wrench



3

1. Application
The P211A,Pediatric Tube Clamp Assembly , 22mm is to be exclusively used for the prosthetic fitting of lower extremity amputations.

2. Description
The Tube Clamp Assembly, 22mm is used to connect any component with 22mm tube system. The four set screws are used for anchoring the mating pyramid system as well as for making angular adjustment for alignment. Apply recommended torque to alignment adjustment screws for connection. Apply locate to all screws once the alignment and connection is finalized.

3. Warranty
Warranty applies only when the product is used as intended without modification, following all manufacturer's recommendations. ST&G offers a 2 year warranty on this product.

Store at room temperature, and avoid extremely hot, cold, or moist storage environments

Storage Temperature Range: between temperatures of -20°C to 80°C [-4°F and 176°F]

Do not store in wet or very hot environment



2

5. Declaration of Conformity
ST&G as manufacturer declares that the The P211A,Pediatric Tube Clamp Assembly , 22mm conforms to the requirements of Council EU 2017/745 (EU), ST&G products are manufactured and tested for a particular weight and activity impact level. Use by another user for whom it was not originally manufactured may cause injury and voids any written or implied warranty, or liability.

6. Useful Life
Service life of the product is covered by the warranty period.

7. Disposal
The device and its packaging must be disposed of in accordance with respective national/local environmental regulations.

Notice

Tighten all the screws, which are side by side, using the torque wrench



3

1. Application
The P211A,Pediatric Tube Clamp Assembly , 22mm is to be exclusively used for the prosthetic fitting of lower extremity amputations.

2. Description
The Tube Clamp Assembly, 22mm is used to connect any component with 22mm tube system. The four set screws are used for anchoring the mating pyramid system as well as for making angular adjustment for alignment. Apply recommended torque to adjustment screws for connection. Apply locate to all screws once the alignment and connection is finalized.

3. Warranty
Warranty applies only when the product is used as intended without modification, following all manufacturer's recommendations. ST&G offers a 2 year warranty on this product.

Store at room temperature, and avoid extremely hot, cold, or moist storage environments

Storage Temperature Range: between temperatures of -20°C to 80°C [-4°F and 176°F]

Do not store in wet or very hot environment



2

5. Declaration of Conformity
ST&G as manufacturer declares that the The P211A,Pediatric Tube Clamp Assembly , 22mm conforms to the requirements of Council EU 2017/745 (EU), ST&G products are manufactured and tested for a particular weight and activity impact level. Use by another user for whom it was not originally manufactured may cause injury and voids any written or implied warranty, or liability.

6. Useful Life
Service life of the product is covered by the warranty period.

7. Disposal
The device and its packaging must be disposed of in accordance with respective national/local environmental regulations.

Notice

Tighten all the screws, which are side by side, using the torque wrench



3