



# BK Kit

## Instructions For Use

110



110A



110S



110T

### 1. Application & Description

- These kits consist of Pylon with Receiver (**1211**), and Tube Clamp (**1212**) in Aluminum, Stainless Steel, or Titanium, and is exclusively for lower extremity prosthetic fittings.
- Connects any pyramid system to another pyramid system.
- The four set screws anchor the mating system and allow for angular alignment.
- The Tube Clamp screw can be loosened to allow for rotation adjustment.

### 2. Technical Specifications

| P/N  | Material        | System Height | Weight | Weight Limit |
|------|-----------------|---------------|--------|--------------|
| 110A | Aluminum        | 415mm         | 390g   | 158kg/350lbs |
| 110S | Stainless Steel | 415mm         | 472g   | 136kg/300lbs |
| 110T | Titanium        | 415mm         | 385g   | 170kg/375lbs |

- **Recommended Torque:** Always cross tighten Alignment Set Screws
- **M8 Alignment Set Screw (1212 & 1211):**
  - (Aluminum): **10Nm**
  - (Titanium/Stainless Steel): **15Nm**
- **M6 Cap Screw (1212 only):**
  - (Stainless Steel): **7Nm**
- **M5 Cap Screw (1212 only):**
  - (Titanium): **7Nm**
  - (Aluminum): **5Nm**
- **Thread Locker:** Apply Loctite once alignment and connection are finalized

### 3. Safety & Environment

- **Temperature Range:** Between -20°C to 80°C (-4°F to 176°F). 
- Rinse with clean water if exposed to salty, chlorinated, or otherwise potentially corrosive liquids.
- **Storage:** Store at room temperature; avoid wet, very hot, or extremely cold environments
- **Warranty:** 2 years, provided it is used as intended without modification, and only for the original intended user.
- Do not reuse on another prosthesis. 
- **Conformity:** This product meets the requirements Council EU 2017/745 (MDR) for medical products. This product has been classified as a class I product according to the classification criteria outlined in appendix IX of the guidelines. Please keep this manual in safe place for future reference.

110 IFU REV. B (2-27-2026)  
Specifications subject to change.