



1121

Suggested:
L5972

The PRIME PLUS Foot provides K2 users a cost-effective, yet highly functional multi-axial foot with a variety of features utilizing a high-tech double keel composite construction and integrated foot shell. The PRIME PLUS is design for spirited walking that has enhanced heel compression for increased foot flat at loading response to mid-stance, allowing for ease of single limb support, walking stability, and smooth dynamic roll over into swing due to the curved dual dynamic keels. It is the evolution of the Prime for the Active K2 user!

Features and Benefits

- Low profile
- Dual High Tech composite curved keel for optimized movement
- Compliant heel and split keel helps initiate foot flat to increase walking stability on uneven terrain providing multi-axial stability.
- foot shell and spectra sock included
- Integrated pyramid
- Split Toe Design
- Water Proof design foot shell drain hole

Size	Cat 1	Cat 2	Cat 3	Cat 4	Cat 5	Cat 6
22						
23						
24						
25						
26						
27						
28						
29						
max wt	50kg	70kg	90kg	100kg	125kg	150kg

** Check the Foot Matrix for weight limits and available sizes by category.

FOOT INFORMATION	
Product #	1121
Sizes:	22-29
Weight of Foot Size 26	620g
Build Height of Foot Size 26	73mm
Heel Height:	10mm
Maximum Patient Weight**	150kg
Warranty	2 year

Note: When placing an order, please use the below formula.

1121-size L or R

Example:

1121-C3-25R (For Beige Foot Shell)

1121-B-C3-25R (For Brown Foot Shell)

- Foot shell comes in beige or brown
- Replacement available



P/N	Description
1121-SH-XXL/R	Foot Shell size 22-29 L/R Beige
1121-B-SH-XXL/R	Foot Shell size 22-29 L/R Brown

P/N	Description
4200	Foot Shell Removal Tool



INTENDED USE

The Prime is a prosthetic foot designed to replace one or more functions of the biologic human foot. It is recommended to use ST&G components with the Prime foot

INDICATIONS:

- K2 amputees with a lower limb amputations

PROTECTIVE COVER ON PYRAMID DOME:

Remove the protective cover on the pyramid dome after alignment is completed and before patient leaves clinic. This protective cover is to eliminate pyramid gouging in case foot needs to be returned for any reason prior to delivery of the prosthesis. Otherwise, a returned foot with damage will incur a fee.

TECHNICAL SPECIFICATIONS


FOOT SIZE	WEIGHT LIMIT**	BUILD HEIGHT	FOOT WEIGHT*
22-23 cm	154lbs / 70 kg	2.76in / 70mm	
24cm	198lbs / 90kg	2.76in / 70mm	
25cm	120lbs / 100kg	2.87in / 73mm	
26-27 cm	275lbs / 125 kg	2.87in / 73mm	620 g
28-29 cm	330lbs / 150 kg	2.99in / 76mm	

*26cm Complete foot with foot shell and sock

** Check the Foot Matrix for weight limits and available sizes by category.

MOUNTING

Use only high quality proximal endoskeletal components. The service-free Prime foot comes assembled in a foot shell within a Spectra Sock.



STATIC ALIGNMENT

For optimal function, balance the patient's weight evenly between the heel and toe.

- The Prime was designed with a 3/8" (10 mm) heel rise.
- The load line divides the foot at the rear 1/3 of the foot

DYNAMIC ALIGNMENT

For a firmer toe response, plantarflex the foot, or move load line posterior

For a softer toe response, dorsiflex the foot, or move load line anterior

For a firmer heel response, dorsiflex the foot, or move load line anterior

For a softer heel response, plantarflex the foot, or move load line posterior

WARNING

- Do not expose this product to pH extremes or corrosive materials (I.E salt water)
- Failure to follow these technical instructions or use of this product outside the scope of the Warranty may result in injury to the patient or damage to the product, and will void the warranty
- Any modification of components will void the warranty.

WARRANTY INSPECTION / MAINTENANCE INFORMATION

It is recommended that you schedule your patients for check-ups according to the Warranty Inspection schedule below. Ensure foot shell and Spectra sock are clean, and weep holes are clear.

High patient weight and/or impact level may require more frequent inspections. We recommend you visually inspect the following applicable parts for excessive wear and fatigue Bi-Annually.

- Composite keel and Adapter
- Foot Shell
- Spectra Sock

- Never allow aggregates such as sand to remain in the foot shell. Upon exposure to aggregates, immediately disassemble foot module and Spectra™ sock and thoroughly rinse out aggregate with water. The abrasive properties of aggregates will quickly wear the graphite components of the foot module.
- Foot modules are manufactured to fit industry standard pyramids and receivers. It is the prosthetists' responsibility to properly select fitting attachment components.
- Never attempt to loosen the bolt affixing the pyramid connector.
- Discontinue use and consult your prosthetist if any part of the prosthesis starts to make noise, or has reduced performance. This may be a sign of abnormal wear, or possible failure.
- Inform your prosthetist if you lose or gain a significant amount of weight.
- ST&G Foot products are manufactured and tested for a particular weight and activity impact level. Use by another user for whom it was not originally manufactured may cause injury and shall void any written or implied warranty.

Foot shell and Spectra sock should not be exposed to extreme and continuous UV exposure, as it will prematurely fade and break down the material. Always use a covering such as a clean sock and shoes, and be sure to routinely change the sock so as to keep the enclosed environment clean and debris free, do not walk bare footed, as this will void warranty.

STORAGE / ENVIRONMENT

Operating & Storage Temperature Range: -10°C to 50°C (14°F to 122°F)

Allowable relative humidity 0 % to 90 %, non-condensing

Can be submerged up to 1 meter.

Rinsed with clean water when exposed to chlorinated pool water.

Do not expose to Salt Water, this will void warranty!



LIABILITY

The manufacturer recommends using the device only under the specified conditions and for the intended purposes. The device must be maintained according to the instructions for use. The manufacturer is not liable for damage caused by component combinations that were not authorized by the manufacturer.

Always apply Blue Loctite to all alignment screws and assembly fasteners with appropriate torque. This is the responsibility of the Certified Prosthetist to ensure device integrity. Do not over torque the alignment screws, as this will damage the foot mounting pyramid faces.



CE Conformity

This product meets the requirements Council EU 2017/745 (MDR) for medical products. This product has been classified as a class I product according to the classification criteria outlined in appendix IX of the guidelines. Please keep this manual in safe place for future use.

Warranty covers defects in workmanship for the given warranty period. Signs of abuse, lack of biannual inspection, improper installation are examples of areas not covered by warranty.

Useful Life

Service life of the product is covered by the warranty period. Not to be used on another wearer other than original intended wearer.



Disposal

The device and its packaging must be disposed of in accordance with respective national/local environmental regulations.

Do not expose Carbon feet to sand, talcum powder, or wet environments

Warranty can be void due to: Unusual usage or abuse (ex. excessive dusty/sand environmental, salt water exposure, high impact (jogging, running, jumping), non-typical usage), Tampering with any fasteners, Modification to the module such as sanding or damaging the finish, patient weight too close to weight limit of foot module (we require a 20-25 lb. buffer).

Warranty Period is 3 years for the Keel and 6 months for the foot shell.

* Alignment screw galling will incur charges if trying to return foot. Utilize the pyramid protector during initial alignment, or request a demo foot!

1121 IFU REV. E (7-6-26)

(Specifications subject to change)